

[Home](#)[Up](#)[Contents](#)

# Colorado Plateau Stable Isotope Laboratory

---

## Sample submission - organics - $\delta^{13}\text{C}$ and $\delta^{15}\text{N}$ via EA-CFIRMS

**Please note: the information on this page is available for download, by clicking [here](#).**

**Our sample-submission sheet (for prepared samples) is available, by clicking [here](#) (pdf version), or [here](#) (MS Word version).**

### Drying, grinding, and weighing

Samples must be dried, ground, and weighed prior to isotope analysis.

**Please note** that soils should be sieved to a consistent particle size prior to grinding and weighing. **Also**, soils containing carbonate materials may exhibit erroneous (organic)  $\delta^{13}\text{C}$  values unless acid-treatment is performed prior to analysis.

Drying can be accomplished via a drying oven at 50-60°C for 24-48 h, or via a freeze drier. Grinding can be achieved via a mortar and pestle, a ball-mill grinder, a mixer mill, etc. The goal is to improve sample homogeneity (i.e., isotopic homogeneity) by way of pulverizing the sample into a fine powder or flour. After drying and grinding, samples need to be weighed into small tin capsules prior to isotope analysis. Capsule size will depend on sample weight. For example, 4x6-mm tin capsules are preferred when sample weights are less than 7 mg; 5x9-mm tin capsules are used for sample weights between 7 mg and 40 mg; and 9x10-mm tin capsules are used for samples greater than 40 mg. Tin capsules can be purchased from [COSTECH Analytical Inc.](#) (800-524-7219), [CE Elantech](#) (888-232-4676), [Elemental Microanalysis](#) (978-526-8517), or in Canada from [Isomass Scientific](#) (800-363-7823), or from any other company supplying products and services for isotope-ratio and/or elemental analysis.

Using a micro-analytical balance, the mass of a sample should be determined to 3 decimal places on a milligram. For example, you should be able to record the mass of a sample as "2.104 mg". Please note that if you are unable to acquire a balance that measures samples to 3 decimal places on a milligram, but instead can only use a balance that measures to 2 decimal places on a milligram (e.g., 2.10 mg) or to 1 decimal place on a milligram (e.g., 2.1 mg), it is important to realize that some precision will be lost when measuring %C and %N (but not C/N). This is because %C and %N are calculated using the recorded mass of the sample. Isotope measurements are not affected by small inaccuracies in weighing because their determination is not mass-dependent (beyond the sole assurance that a large enough sample has been combusted in the EA). Thus if one can live with slightly lower precision on % C and %N (e.g.,  $\pm 3.0\%$  for carbon as opposed to  $\pm 0.5\%$ ) then acquiring a balance that records the mass of the sample to 3 decimal places on a milligram is not essential.

Prior to weighing out the sample, do not forget to tare the balance, in order to exclude the initial mass of the tin capsule. Once the appropriate amount of sample has been placed in the tin capsule, the capsule needs to be "crushed" into a small ball (or square). This can be done by gently applying force using a pair of forceps, or by rolling the sample gently between your thumb and index finger. At no point in time should you touch the sample or the tin capsule with your bare hands. Powder-less latex gloves should be used if you plan on "crushing" the capsule between your fingers. I usually use 2 pairs of forceps to handle the sample, so that I never have to touch it with my bare hands. Please note that if a capsule containing a sample falls on the floor, it should be thrown away. Also, please be sure that

material will not leak from the capsule after it has been crushed, as the loss of material can affect both the isotope and elemental data (through incorrect mass determination and subsequent sample-to-sample contamination). Once the tin capsule has been crushed, please re-weigh the sample to confirm and record the final mass.

[back to top](#)

## Mass requirements

The required (dry) mass for simultaneous analysis of  $\delta^{15}\text{N}$  and  $\delta^{13}\text{C}$  depends mostly on the sample type, but also on the nitrogen content of each sample. Since nitrogen occurs in lower concentrations than carbon (for most organic materials), the nitrogen content becomes the limiting factor in dual-isotope measurements. Plants (i.e., 2-5% nitrogen) and soils/sediments (0.1-1.0% nitrogen) have much lower nitrogen concentrations than animals (i.e., 8-12% nitrogen), so more material is needed per sample for plants and soils than for animals. ***The preference in our isotope lab is to weigh out 60 micrograms of nitrogen per sample (regardless of the sample type).*** This amount of material usually results in a nitrogen peak of sufficient size for accurately measuring stable isotopes, but it does not result in a peak so large that it will exhaust the chemicals too quickly. Since there is usually much more carbon than nitrogen in organic materials, there should always be enough carbon per sample for  $\delta^{13}\text{C}$  analysis (so long as samples have been properly prepared for  $\delta^{15}\text{N}$  analysis).

**Please note: there is both a LOWER limit and an UPPER limit to the amount of sample that should be submitted in each tin cup. Please follow the guidelines below for the different sample types, and if you have any questions, do not hesitate to contact the [lab manager](#) for more information.**

To achieve 60 micrograms of nitrogen per sample, one needs to know the approximate nitrogen content of the material being analyzed. As mentioned above, plants usually contain less nitrogen than animals, so to obtain 60 micrograms of N for plants with a nitrogen content of approximately 2%, one would need to weigh out approximately 3.000 milligrams of dried plant material. For animals (e.g., fish, birds, mammals, invertebrates, etc.) with a nitrogen content of approximately 10%, one would need to weigh out approximately 0.600 milligrams of dried material to obtain 60 micrograms of N. For ease of measurement, we usually ask clients to weigh out plant material between 3.000 and 4.000 mg. For animal material, we ask that you weigh out samples between 0.800 mg and 1.200 mg.

Below, there is a small table outlining required sample masses, given the approximate % nitrogen content of different sample materials:

Milligrams of dried material to be weighed per sample	%N of sample	Micrograms of N needed per sample for isotope analysis	Possible sample types
60.000	0.1%	60	soil
12.000	0.5%	60	soil
4.000	1.5%	60	decaying plant litter
2.000	3.0%	60	fresh leaves, algae
1.200	5.0%	60	legumes
0.600	10.0%	60	all animals
0.300	20.0%	60	ammonium salts
0.150	40.0%	60	urea

If you are unsure of the nitrogen content of your sample (e.g., bacteria, soils, sediments, FPOM, etc.), it would be best to send a preliminary sample of the material (in a vial) to our lab so that we could analyze it for you and recommend an appropriate mass. Another alternative is for you to send ALL of your samples to our lab to be weighed here, at which point we would be able to run repeats for you on larger/smaller samples if necessary.

Once weighed, crushed, and re-weighed, samples should be placed into a 96-well cell plate (with cover) for easy storage and shipping. We use 96-well polystyrene plates. [Cell plates](#) can be ordered (in bulk) from companies such as [Fisher-Scientific Inc](#) or [VWR International](#). For smaller quantities,

plates can be purchased from [COSTECH Analytical Inc.](#) Each well in the cell plate has an alpha-numeric position (rows A through H, columns 1 through 12). Thus, the first well is A1, the next A2, and so on. Once you are ready to weigh out your samples, please download our [sample-submission sheet](#) so that you can provide us with enough information to successfully analyze your samples.

[back to top](#)

## Cost of simultaneous $\delta^{13}\text{C}$ and $\delta^{15}\text{N}$ analyses via EA-CFIRMS

Sample costs are:

\$9 per sample (if **you** grind and **you** weigh)

\$12 per sample (if **you** grind and **we** weigh)

\$15 per sample (if **we** grind and **we** weigh)

For this price, you get  $\delta^{15}\text{N}$ ,  $\delta^{13}\text{C}$ , %C, %N, and C/N for each sample.

An additional \$3 per sample will be included for any other tasks that need to be performed by our lab prior to isotope analysis (e.g., sieving soils, pre-combustion of sediments to remove organics, acid-treatment to remove inorganic carbon, etc.). Costs are reduced to a minimum of \$7.50 per sample if samples are submitted in bulk ( $n > 500$ ).

**For additional prices on other analyses, click [here](#).**

[back to top](#)

## Turn-around Time

Once the samples arrive in our lab, we will confirm receipt of the shipment via email. Turn-around time in our lab depends mostly on the state of preparedness of your samples. For example, if your samples arrive weighed out and crushed into small tin cups, then analysis should be complete within 4 weeks of our receiving the shipment (depending on the size of the shipment). However, if the samples need to be prepped in our lab before analysis (e.g., ground, weighed, acid-treated, etc.), then the turn-around time could be on the order of 4-10 weeks. **Please note:** we are currently in the process of purchasing a second EA. This second EA-IRMS system will allow us to run twice as many samples per day, cutting turn-around times in half.

Once the samples have been analyzed, we will send you a copy of the data via email. A hardcopy will follow by regular mail along with the invoice for services. Included in the data file will be the results of quality assurance and quality control (QA/QC). We use a NIST standard (peach leaves) and four IAEA standards (IAEA-N1, N2, CH6, and CH7) to correct the raw isotope data. We also use a series of five elemental standards to correct the %C and the %N data. Standards are interspersed throughout the run, to ensure data integrity over the entire analysis. As you will note on the sample-submission sheet, we ask that you weigh out duplicates of some samples as a check on sample homogeneity. If for some unknown reason (e.g., power failure, exhaustion of chemicals, computer error) a sample is lost during analysis, we will request that you re-submit the material. Any repeats that are requested by the lab will be analyzed free of charge.

[back to top](#)

## Shipping

**To send samples to our lab, please click [here](#) for our shipping address.**

## Additional Information

We run samples for  $\delta^{15}\text{N}$ ,  $\delta^{13}\text{C}$ , %C, %N, and C/N, in continuous-flow mode using a Thermo-Finnigan Delta<sup>plus</sup> Advantage gas isotope-ratio mass spectrometer interfaced with a Carlo Erba 2100 elemental

analyzer. Helium flow rate is usually set at 130 ml/min. Oxygen flow rate is at 80 ml/min. A standard 4-meter GC column is used (set at 65°C) for peak separation, in combination with one quartz (combustion) tube filled with chromium oxide and silvered cobaltous/cobaltic oxide (set at 1000°C) and one quartz (reduction) tube filled with reduced copper (set at 780°C). Dual analysis of  $\delta^{15}\text{N}$  and  $\delta^{13}\text{C}$  usually takes about 10 minutes per sample. The average daily run contains approximately 125 samples (incl. standards). Data are normalized using 4 internationally-accepted isotope standards (IAEA CH6, CH7, N1, and N2). Our working standard is peach leaves (NIST 1547). External precision on these standards is  $\pm 0.10\text{‰}$  or better for  $\delta^{13}\text{C}$  and  $\pm 0.20\text{‰}$  or better for  $\delta^{15}\text{N}$ .  $\delta^{13}\text{C}$  and  $\delta^{15}\text{N}$  data are expressed relative to VPDB for carbon, and to AIR for nitrogen.

[back to top](#)

[CPSIL](#)

[ [Home](#) ] [ [Up](#) ]

**Acknowledgements:** CPSIL has been in operation since May, 1999, and was funded through a grant from the National Science Foundation (DBI 9729600) and from Northern Arizona University.

Send mail to [Richard.Doucett@nau.edu](mailto:Richard.Doucett@nau.edu) with questions, comments, and/or corrections about this web site.  
Last modified: 01/15/04. This site is best viewed using Microsoft Internet Explorer 6.0.